



STS

Specializing in Laboratory Services, Sterilization & Packaging

7500 W. Henrietta Road, Rush, NY 14543 Phone: 585.533.1672 Fax: 585.533.1796

USP Class Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attn: Matt Tones

Test Nos.: T04-2417, 2418, 2419

Test Material: GYLON® Style 3504 (blue)

This test material meets the requirements for USP Class VI designation at a temperature extraction of 121°C for 1 hour.

Jonia Schaubroeck

Study Director

7/14/04

Date